

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0025
Location:	CLIFTON CITY POLICE	
Calibration File No.:	00904	Calib. Date: 09/03/2009
Certification File No.:	00905	Calib. No.: 00017
Linearity File No.:	00906	Cert. Date: 09/03/2009
Solution File No.:	01021	Cert. No.: 00010
Sequential File No.:	01021	Lin. Date: 09/03/2009
		Lin. No.: 00010
		Soln. Date: 12/24/2009
		Soln. No.: 00084
		File Date: 12/24/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0033
Solution Control Lot:	09A06Z	Expires: 01/15/2011
		Bottle No.: 1181

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:29S		
Control 1 EC	0.104%	14:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	14:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30S		
Control 2 EC	0.103%	14:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32S		
Control 3 EC	0.102%	14:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	14:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BRANCATO

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 10

Date: 12/24/2009



State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
Post Office Box 7068  
West Trenton NJ 08628-0068  
(609) 882-2000

JON S. CORZINE  
Governor

ANNE MILGRAM  
Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 2/11/2009

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 09A062

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1195 to 0.1197 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2011.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

*Ajit R. Tungare*  
Ajit R. Tungare  
Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 30<sup>th</sup> day of February, 2009.


*Linda L. DeSantis*

Linda L. DeSantis  
My Commission  
Expires Aug. 17, 2009



DEPARTMENT OF

# State Highway Safety



CLIFTON CHIEF POLICE DEPARTMENT

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS

DAY OF

MARCH

TWO THOUSAND AND SIXTY

NEW JERSEY STATE POLICE

STATE OF NEW JERSEY

*Joseph A. ...*  
SUPERINTENDENT

*William V. ...*  
ATTORNEY GENERAL

DEPARTMENT OF  
**Law and Public Safety**  
*Where the world begins*

**David Brancato**  
 Clifton Police Department

IN QUALIFICATION FOR THE COURSE OF THE STATE POLICE TRAINING CENTER OF  
 THE NEW JERSEY STATE POLICE IN THE COURSE OF THE STATE POLICE TRAINING CENTER OF  
 THE NEW JERSEY STATE POLICE IN THE COURSE OF THE STATE POLICE TRAINING CENTER OF

A METHOD TO DETERMINE INSTRUCTIONS  
 GIVEN UNDER HAND AT TRAINING CENTER IN THE COURSE OF THE STATE POLICE TRAINING CENTER OF

FRONTMONT, N.J. 07034

DATE: February

*[Signature]*  
 SUPERVISOR  
 NEW JERSEY STATE POLICE

*[Signature]*  
 SUPERVISOR  
 NEW JERSEY STATE POLICE

ORIGINAL COURSE DATES 3-29-2006

DATE	Refresh Course PLACE	INSTRUCTOR
1. <u>2-7-08</u>	<u>Sussex Co. PA.</u>	<u>Russo</u>
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2333 (Rev. 07/07)