

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: CLIFTON CITY POLICE
Calibration File No.: 00904
Certification File No.: 00905
Linearity File No.: 00906
Solution File No.: 01021
Sequential File No.: 01021

Serial No.: ARWM-0025

Calib. Date: 09/03/2009
Cert. Date: 09/03/2009
Lin. Date: 09/03/2009
Soln. Date: 12/24/2009
File Date: 12/24/2009

Calib. No.: 00017
Cert. No.: 00010
Lin. No.: 00010
Soln. No.: 00084

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 09A062

Model No.: CU-34

Serial No.: DDXA S3-0033
Expires: 01/15/2011
Bottle No.: 1181

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:29S		
Control 1 EC	0.104%	14:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	14:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30S		
Control 2 EC	0.103%	14:31S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32S		
Control 3 EC	0.102%	14:32S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	14:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRANCATO

First Name: DAVID

MI: A

Signature: 

Badge No.: 10

Date: 12/24/2009



State of New Jersey
OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
Post Office Box 7068
West Trenton NJ 08628-0068
(609) 882-2000

JON S. CORZINE
Governor

ANNE MILGRAM
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 2/11/2009

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 09A062

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1195 to 0.1197 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2011.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
Ajit R. Tungare
Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 30th day of February, 2009.


Linda L. DeSantis

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009



DEPARTMENT OF

Ita In
Summ **Public Safety**



CLIFTON CHIEF POLICE DEPARTMENT

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AND ATTENTION NEW JERSEY THIS 29th DAY OF MARCH

TWO THOUSAND AND 87X

Joseph A. ...
SUPERINTENDENT
NEW JERSEY STATE POLICE

William V. ...
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Law and Public Safety
Officio in Civitate

David Francato
 Clifton Police Department

IN QUALITY AND COMPETENCY TO COMPLY WITH THE REQUIREMENTS OF THE NEW JERSEY CHARTER OF
 THE LAW ENFORCEMENT OCCASION OF THE **ANGOTAS/AUGER INC**

A METHOD TO DETERMINE INSTRUCTION
 GIVEN UNDER HAND AT TRAINING IN **2006** DATE OF **February**

TRAINING AND RIGHT

[Signature]
 SUPERVISOR
 NEW JERSEY STATE POLICE

[Signature]
 SUPERVISOR
 STATE POLICE

ORIGINAL COURSE DATES 3-29-2006

DATE	Refresh Course PLACE	INSTRUCTOR
1. <u>2-7-08</u>	<u>Sussex Co. PA.</u>	<u>Russo</u>
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2333 (Rev. 07/07)