

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0025
Location:	CLIFTON CITY POLICE	
Calibration File No.:	00768	Calib. Date: 03/20/2009
Certification File No.:	00769	Cert. No.: 00016
Linearity File No.:	00770	Cert. Date: 03/20/2009
Solution File No.:	00850	Lin. No.: 00009
Sequential File No.:	00850	Lin. Date: 03/20/2009
		Soln. No.: 00072
		Soln. Date: 06/17/2009
		File Date: 06/17/2009

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0032
Solution Control Lot:	07J048	Expires: 10/03/2009
		Bottle No.: 1068

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:58D		
Control 1 EC	0.099%	09:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:00D		
Control 2 EC	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:01D		
Control 3 EC	0.100%	10:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRANCATO

First Name: DAVID

MI: A

Signature: _____



Badge No.: 10

Date: 06/17/2009



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY

DIVISION OF STATE POLICE

POST OFFICE BOX 7068

WEST TRENTON NJ 08628-0068

(609) 882-2000

ANNE MILGRAM

Attorney General

COLONEL JOSEPH R. FUENTES

Superintendent

JON S. CORZINE

Governor

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 11/28/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07J048

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1225 to 0.1229 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is October 3, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 17th day of December, 2007.

Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009



New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



DEPARTMENT OF

Justice and Public Safety

CLIFTON CITY POLICE DEPARTMENT


IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS

TWO THOUSAND AND SIX

DAY OF

MARCH



Office to certify that

DAVID A. BRANCAIO

Joseph A. Blawie
SUPERINTENDENT
NEW JERSEY STATE POLICE

Roberta V. Parker
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

OF THE STATE OF NEW JERSEY
David Brancato
 Clifton Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT DRUG ALCOHOL TESTS PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1984 IN THE OPERATION OF THE **ALCOHOL 7110-MK III-C**
 A METHOD TO DETERMINE INTOXICATION
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 20th DAY OF February

TWO THOUSAND AND Eight

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 3-29-2006

	DATE	Refresher Course- PLACE	INSTRUCTOR
1.	<u>2-7-08</u>	<u>Sussex Co. PA</u>	<u>Russo</u>
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			