

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWM-0025  
Location: CLIFTON CITY POLICE  
Calibration File No.: 01040 Calib. Date: 02/15/2010 Calib. No.: 00020  
Certification File No.: 01041 Cert. Date: 02/15/2010 Cert. No.: 00011  
Linearity File No.: 01042 Lin. Date: 02/15/2010 Lin. No.: 00011  
Solution File No.: 01057 Soln. Date: 03/15/2010 Soln. No.: 00090  
Sequential File No.: 01057 File Date: 03/15/2010

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0032  
Control Solution %: 0.100% Expires: 01/15/2011  
Solution Control Lot: 09A062 Bottle No.: 1185

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:04S		
Control 1 EC	0.101%	17:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:05S		
Control 2 EC	0.099%	17:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:07S		
Control 3 EC	0.100%	17:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BRANCATO

First Name: DAVID

MI: A

Signature: \_\_\_\_\_

Badge No.: 10

Date: 03/15/2010



JON S. CORZINE  
Governor

State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
Post Office Box 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

ANNE MILGRAM  
Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 2/11/2009

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 09A062

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1195 to 0.1197 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2011.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare  
Chief Forensic Scientist  
Division of State Police


Sworn to and subscribed before me this 30<sup>th</sup> day of February, 2009.

Linda L. DeSantis  
My Commission  
Expires Aug. 17, 2009



DEPARTMENT OF

*John J. Hughes*  
*and*  
*Richard S. Stephens*



This is to certify that  
DAVID A. BRANCATO  
CLIFTON CITY POLICE DEPARTMENT

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO  
CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C  
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29TH DAY OF MARCH

TWO THOUSAND AND SIX

*Joseph P. Linn*  
SUPERINTENDENT  
NEW JERSEY STATE POLICE

*John V. Funder*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

DEPARTMENT OF  
**Law and Public Safety**  
 This is to certify that

OF THE STATE OF NEW JERSEY  
**David Brancato**  
 Clifton Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1966 IN THE OPERATION OF THE ALCOHOL TESTING MK III C  
 A METHOD TO DETERMINE INTOXICATION  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, ON

20th DAY OF February  
 TWO THOUSAND AND EIGHT

  
 SUPERINTENDENT  
 N.J. POLICE STATE POLICE

  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 3-29-2006

DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>2-7-08</u>	<u>Sussex Co. FA</u>	<u>Russo</u>
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____