

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0025
Location: CLIFTON CITY POLICE
Calibration File No.: 00768 Calib. Date: 03/20/2009 Calib. No.: 00016
Certification File No.: 00769 Cert. Date: 03/20/2009 Cert. No.: 00009
Linearity File No.: 00770 Lin. Date: 03/20/2009 Lin. No.: 00009
Solution File No.: 00816 Soln. Date: 05/15/2009 Soln. No.: 00070
Sequential File No.: 00816 File Date: 05/15/2009

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0032
Control Solution %: 0.100% Expires: 10/03/2009
Solution Control Lot: 07J048 Bottle No.: 1066

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:15D		
Control 1 EC	0.099%	15:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:16D		
Control 2 EC	0.099%	15:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:18D		
Control 3 EC	0.100%	15:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:19D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BRANCATO

First Name: DAVID

MI: A

Signature: 

Badge No.: 10

Date: 05/15/2009



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL

DEPARTMENT OF LAW AND PUBLIC SAFETY

DIVISION OF STATE POLICE

POST OFFICE BOX 7068

WEST TRENTON NJ 08628-0068

(609) 882-2000

ANNE MILGRAM

Attorney General

COLONEL JOSEPH R. FUENTES

Superintendent

JON S. CORZINE

Governor

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 11/28/07

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 07J048

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1225 to 0.1229 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is October 3, 2009.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare
Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 17th day of December, 2007.

Notary

Linda L. DeSantis
My Commission
Expires Aug. 17, 2009




New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



DEPARTMENT OF

Tran and Harkler Safety





This is to certify that
DAVID A. BRANCAIO
CLIFTON CITY POLICE DEPARTMENT

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO
CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29TH DAY OF MARCH

TWO THOUSAND AND SIX


SUPERINTENDENT
NEW JERSEY STATE POLICE



ATTORNEY GENERAL
STATE OF NEW JERSEY


DEPARTMENT OF
Law and Public Safety
This is to certify that

David Brancato
Clifton Police Department

IS QUALIFIED AND COMPETENT TO COMPLETE CHRYSLER BREATHANALYZER PURSUANT TO CHAPTER 142 OF THE LAWS OF 1964 IN THE OPERATION OF THE **ALCOHOL/7410 MK III-C**

A METHOD TO DETERMINE INTOXICATION
GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY ON 20th DAY OF February
TWO THOUSAND AND Eight


SUPERINTENDENT
NEW JERSEY STATE POLICE


ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES 3-29-2006

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>2-7-08</u>	<u>Sussex Co. PA</u>	<u>Russo</u>
2.	_____	_____	_____
3.	_____	_____	_____
4.	_____	_____	_____
5.	_____	_____	_____
6.	_____	_____	_____
7.	_____	_____	_____
8.	_____	_____	_____
9.	_____	_____	_____

S.P. 293B (Rev. 07/07)