

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0025
Location:	CLIFTON CITY POLICE	
Calibration File No.:	01160	Calib. Date: 08/12/2010
Certification File No.:	01161	Calib. No.: 00021
Linearity File No.:	01162	Cert. Date: 08/12/2010
Solution File No.:	01246	Cert. No.: 00012
Sequential File No.:	01246	Lin. Date: 08/12/2010
		Lin. No.: 00012
		Soln. Date: 11/21/2010
		Soln. No.: 00105
		File Date: 11/21/2010

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0032
Solution Control Lot:	10C077	Expires: 03/10/2012
		Bottle No.: 0376

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:10S		
Control 1 EC	0.104%	17:11S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	17:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:12S		
Control 2 EC	0.101%	17:12S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	17:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:13S		
Control 3 EC	0.101%	17:14S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	17:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:14S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BRANCATO

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 10

Date: 11/21/2010



State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

CHRIS CHRISTIE  
Governor

KIM GUADAGNO  
Lt. Governor

PAULA T. DOW  
Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.


MANUFACTURER: Drager Safety, Inc. ANALYSIS DATE: 3/24/2010

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10C977


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1198 to 0.1202 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 10, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
Kenneth W. Kawalek, M.S.  
Assistant Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 6<sup>th</sup> day of April, 2010.

  
Notary




New Jersey Is An Equal Opportunity Employer  
Printed on Recycled Paper and Recyclable



DEPARTMENT OF

*more* *Justice* *State*



This is to certify that  
DAVID A. BRANCATO

CLIFTON CITY POLICE DEPARTMENT

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO  
CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C  
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29TH DAY OF MARCH

TWO THOUSAND AND SIX

*Joseph P. ...*  
SUPERINTENDENT  
NEW JERSEY STATE POLICE

*Julian V. Fisher*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

DEPARTMENT OF  
*Law and Public Safety*  
 This is to certify that

**David Brancato**  
**Clifton Police Department**

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOTEST 7110 MK III-C**  
 A METHOD TO DETERMINE INTOXICATION,  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 20th DAY OF February

TWO THOUSAND AND Eight

*[Signature]*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 3-29-2006

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>2-7-08</u>	<u>Sussex Co. FA</u>	<u>Russo</u>
2.	<u>8-3-10</u>	<u>B.C.P.A.</u>	<u>R. T.</u>
3.			
4.			
5.			
6.			
7.			
8.			
9.			