

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0025
Location:	CLIFTON CITY POLICE	
Calibration File No.:	00904	Calib. Date: 09/03/2009
Certification File No.:	00905	Calib. No.: 00017
Linearity File No.:	00906	Cert. Date: 09/03/2009
Solution File No.:	00964	Cert. No.: 00010
Sequential File No.:	00964	Lin. Date: 09/03/2009
		Lin. No.: 00010
		Soln. Date: 10/28/2009
		Soln. No.: 00079
		File Date: 10/28/2009
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0033
Solution Control Lot:	09A062	Expires: 01/15/2011
		Bottle No.: 1176

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:21S		
Control 1 EC	0.102%	08:22S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	08:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:23S		
Control 2 EC	0.101%	08:23S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.103%	08:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:24S		
Control 3 EC	0.101%	08:25S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	08:25S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: BRANGATO

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 10

Date: 10/28/2009



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

JON S. CORZINE  
Governor

ANNE MILGRAM  
Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

**ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION:** Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

**MANUFACTURER:** Drager Safety, Inc.

**ANALYSIS DATE:** 2/11/2009

**BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER:** 09A062

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have an ethyl alcohol concentration range of 0.1195 to 0.1197 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is January 15, 2011.

As Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ajit R. Tungare  
Chief Forensic Scientist  
Division of State Police


Sworn to and subscribed before me this 20<sup>th</sup> day of February, 2009.

Linda L. DeSantis  
My Commission  
Expires Aug. 17, 2009



DEPARTMENT OF

# Justice and Public Safety



DAVID A. BRANCATO  
CLIFTON CITY POLICE DEPARTMENT

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C, A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TENTON, NEW JERSEY THIS

TWO THOUSAND AND SIX

DAY OF

MARCH

NEW JERSEY STATE POLICE

STATE OF NEW JERSEY

*Richard V. Parker*  
ATTORNEY GENERAL

*Joseph J. ...*  
SUPERINTENDENT  
NEW JERSEY STATE POLICE

DEPARTMENT OF  
**Law and Public Safety**  
 This is to certify that

David Brancato  
 Clifton Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1964 IN THE OPERATION OF THE  
 A METHOD TO DETERMINE INTOXICATION  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THE 20th DAY OF February

TWO THOUSAND AND EIGHT

*[Signature]*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 3-29-2006

DATE	Refresher Course PLACE	INSTRUCTOR
1. <u>2-7-08</u>	<u>Sussex Co. FA</u>	<u>Russo</u>
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____

S.P. 2938 (Rev. 07/07)