

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0025
Location:	CLIFTON CITY POLICE	
Calibration File No.:	01307	Calib. Date: 02/15/2011
Certification File No.:	01308	Calib. No.: 00022
Linearity File No.:	01309	Cert. Date: 02/15/2011
Solution File No.:	01434	Cert. No.: 00013
Sequential File No.:	01434	Lin. Date: 02/15/2011
		Lin. No.: 00013
		Soln. Date: 08/04/2011
		Soln. No.: 00122
		File Date: 08/04/2011

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0033
Solution Control Lot:	10C077	Expires: 03/10/2012
		Bottle No.: 0393

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:11D		
Control 1 EC	0.100%	10:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:12D		
Control 2 EC	0.099%	10:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:14D		
Control 3 EC	0.100%	10:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: BRANCATO

First Name: DAVID

MI: A

Signature: \_\_\_\_\_



Badge No.: 10

Date: 08/04/2011



State of New Jersey  
OFFICE OF THE ATTORNEY GENERAL  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
POST OFFICE BOX 7068  
WEST TRENTON NJ 08628-0068  
(609) 882-2000

CHRIS CHRISTIE  
Governor

KIM GUADAGNO  
Lt. Governor

PAULA T. DOW  
Attorney General

COLONEL JOSEPH R. FUENTES  
Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Drager Safety, Inc.

ANALYSIS DATE: 3/24/2010


BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 10C077

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1198 to 0.1202 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 10, 2012.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

  
Kenneth W. Kawalek, M.S.  
Assistant Chief Forensic Scientist  
Division of State Police

Sworn to and subscribed before me this 24 day of April, 2010.  
  
Notary

Linda L. Fuentes  
Notary Public, New Jersey  
My Commission Expires 8-17-14




New Jersey Is An Equal Opportunity Employer  
Printed on Recycled Paper and Recyclable



DEPARTMENT OF

*Notar Public*



This is to certify that  
DAVID A. BRANCATO  
CLIFTON CITY POLICE DEPARTMENT

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO  
CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C  
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 29<sup>TH</sup> DAY OF MARCH

TWO THOUSAND AND SIX

*Joseph J. ...*  
SUPERINTENDENT  
NEW JERSEY STATE POLICE

*Julius V. Fisher*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

DEPARTMENT OF  
*Law and Public Safety*  
 This is to certify that

David Brancato

Clifton Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 14 OF THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOTEST 7110 MK III-C** A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, 20th DAY OF February  
 TWO THOUSAND AND Eight

*[Signature]*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES 3-29-2006

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	<u>2-7-08</u>	<u>Sussex Co. FA</u>	<u>Russo</u>
2.	<u>8-3-10</u>	<u>BCPA</u>	<u>A. T.</u>
3.			
4.			
5.			
6.			
7.			
8.			
9.			