

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWM-0025
Location:	CLIFTON CITY POLICE	
Calibration File No.:	01671	Calib. Date: 08/15/2012
Certification File No.:	01672	Calib. No.: 00025
Linearity File No.:	01673	Cert. Date: 08/15/2012
Solution File No.:	01695	Cert. No.: 00016
Sequential File No.:	01695	Lin. Date: 08/15/2012
		Lin. No.: 00016
		Soln. Date: 09/26/2012
		Soln. No.: 00150
		File Date: 09/26/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0033
Solution Control Lot:	11H094	Expires: 08/17/2013
		Bottle No.: 1365

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:30D		
Control 1 EC	0.099%	15:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:32D		
Control 2 EC	0.097%	15:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:33D		
Control 3 EC	0.098%	15:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: TURANO

First Name: PETER

MI: A

Signature: *Peter A. Turano #204*

Badge No.: 204

Date: 09/26/2012



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor

KIM GUADAGNO
Lt. Governor

PAULA T. DOW
Attorney General

COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 9/8/2011

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 11H094

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1197 to 0.1207 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 17, 2013.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Signature]
Kenneth W. Kawalek, M.S.
Assistant Chief Forensic Scientist
Division of State Police

Sworn to and subscribed before me this 12th day of September, 2011.


[Signature]
Notary

Notary Public, New Jersey
My Commission Expires 9-17-14



DEPARTMENT OF

# Motor and Public Safety



This is to certify that

Peter A. Turano  
Clifton City Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS Twelfth DAY OF June

TWO THOUSAND AND Nine

*[Signature]*  
SUPERINTENDENT  
NEW JERSEY STATE POLICE

*[Signature]*  
ATTORNEY GENERAL  
STATE OF NEW JERSEY

DEPARTMENT OF  
**Motor and Public Safety**  
 This is to certify that

**Peter A. Turano**  
 Clifton City Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF  
 THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOTEST 7110 MKIII-C**  
 A METHOD TO DETERMINE INTOXICATION.  
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 12th DAY OF JUNE  
 TWO THOUSAND AND Nine

*[Signature]*  
 SUPERINTENDENT  
 NEW JERSEY STATE POLICE

*[Signature]*  
 ATTORNEY GENERAL  
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES )))

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	11/21/11	BCPA	R. MW
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			