

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0025
Location: CLIFTON CITY POLICE
Calibration File No.: 01671 Calib. Date: 08/15/2012 Calib. No.: 00025
Certification File No.: 01672 Cert. Date: 08/15/2012 Cert. No.: 00016
Linearity File No.: 01673 Lin. Date: 08/15/2012 Lin. No.: 00016
Solution File No.: 01721 Soln. Date: 11/21/2012 Soln. No.: 00154
Sequential File No.: 01721 File Date: 11/21/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0033
Control Solution %: 0.100% Expires: 08/17/2013
Solution Control Lot: 11H094 Bottle No.: 1369

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:12S		
Control 1 EC	0.097%	19:13S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:14S		
Control 2 EC	0.095%	19:14S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	19:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:15S		
Control 3 EC	0.096%	19:16S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	19:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:16S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TURANO

First Name: PETER

MI: A

Signature: 

Badge No.: 204

Date: 11/21/2012



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

PAULA T. DOW
 Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.


ANALYSIS DATE: 9/8/2011

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 11H094


Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1197 to 0.1207 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is August 17, 2013.

As Assistant Chief Forensic Scientist of the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at my direction and under my supervision by personnel of, and at, the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


 Kenneth W. Kawalek, M.S.
 Assistant Chief Forensic Scientist
 Division of State Police

Sworn to and subscribed before me this 12th day of September, 2011.


 Notary

Linda L. DeSantis
 Notary Public, New Jersey
 My Commission Expires 8-17-14




New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable



DEPARTMENT OF

Traffic and Public Safety



This is to certify that

Peter A. Furano
Clifton City Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS Twelfth DAY OF June

TWO THOUSAND AND Nine

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY


DEPARTMENT OF
Motor and Public Safety
 This is to certify that

Peter A. Turano
 Clifton City Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C
 A METHOD TO DETERMINE INTOXICATION.
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 12th DAY OF June

TWO THOUSAND AND Nine


 SUPERINTENDENT
 NEW JERSEY STATE POLICE


 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course	
		PLACE	INSTRUCTOR
1.	11/21/11	BC-PA	R. MW
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			