

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWM-0025
Location:	CLIFTON CITY POLICE	
Calibration File No.:	01823	Calib. Date: 07/03/2013
Certification File No.:	01824	Calib. No.: 00029
Linearity File No.:	01825	Cert. Date: 07/03/2013
Solution File No.:	01851	Cert. No.: 00019
Sequential File No.:	01851	Lin. Date: 07/03/2013
		Lin. No.: 00019
		Soln. Date: 08/29/2013
		Soln. No.: 00171
		File Date: 08/29/2013
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDXA S3-0033
Solution Control Lot:	12K108	Expires: 11/05/2014
		Bottle No.: 0285

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:49D		
Control 1 EC	0.099%	21:50D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:50D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:50D		
Control 2 EC	0.099%	21:51D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:51D		
Control 3 EC	0.099%	21:52D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:53D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TURANO

First Name: PETER

MI: A

Signature: _____

Peter A. Turano

Badge No.: 204

Date: 08/29/2013



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

JEFFREY S. CHIESA
 Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 12/13/2012

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 12K108

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1207 to 0.1218 grams per 100 milliliters of solution.


This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is November 05, 2014.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 21st day of December, 2012.

Linda L. W. Santos
 Notary




"An Internationally Accredited Agency"

New Jersey is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable



DEPARTMENT OF

Tax and Motor Vehicle Safety



This is to certify that
Peter A. Furano
Clifton City Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO
CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS Twelfth DAY OF June

TWO THOUSAND AND Nine

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

Peter A. Turano
 Clifton City Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOTEST 7110 MKIII-C**
 A METHOD TO DETERMINE INTOXICATION.
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 12th DAY OF June

TWO THOUSAND AND Nine

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

1.	DATE	Refresher Course	
		PLACE	INSTRUCTOR
1.	11/21/11	BCPA	R. IV
2.			
3.			
4.			
5.			
6.			
7.			
8.			
9.			