

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0025
Location: CLIFTON CITY POLICE
Calibration File No.: 01917 Calib. Date: 07/02/2014 Calib. No.: 00032
Certification File No.: 01918 Cert. Date: 07/02/2014 Cert. No.: 00021
Linearity File No.: 01919 Lin. Date: 07/02/2014 Lin. No.: 00021
Solution File No.: 01979 Soln. Date: 11/21/2014 Soln. No.: 00182
Sequential File No.: 01979 File Date: 11/21/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0032
Control Solution %: 0.100% Expires: 07/18/2015
Solution Control Lot: 13G115 Bottle No.: 0751

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:41S		
Control 1 EC	0.102%	12:42S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	12:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43S		
Control 2 EC	0.100%	12:43S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:43S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:44S		
Control 3 EC	0.101%	12:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:46S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TURANO

First Name: PETER

MI: A

Signature: 

Badge No.: 204

Date: 11/21/2014



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL

DEPARTMENT OF LAW AND PUBLIC SAFETY

DIVISION OF STATE POLICE

POST OFFICE BOX 7068

WEST TRENTON, NJ 08628-0068

(609) 882-2000

JOHN J. HOFFMAN

Acting Attorney General

COLONEL JOSEPH R. FUENTES

Superintendent

CHRIS CHRISTIE

Governor

KIM GUADAGNO

Lt. Governor

CERTIFICATION OF ANALYSIS

0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 7/29/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 13G115

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1213 to 0.1220 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 18, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.


[Handwritten signature of Ali M. Alaoui]

Ali M. Alaoui, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 7th day of August, 2013.

[Handwritten signature of Linda L. W. Santos]

Notary



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



DEPARTMENT OF
Traffic and Public Safety
 This is to certify that

Peter A. Turano
 Clifton City Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE **ALCO TEST 7110 MKIII-C**
 A METHOD TO DETERMINE INTOXICATION.
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS 12th DAY OF June

TWO THOUSAND AND Nine

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE


[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	Refreshers Course		INSTRUCTOR
	DATE	PLACE	
1.	11/21/11	BC-PA	R. TW
2.	11/13/13	S.C.E.S.T.A.	R. D. J.
3.			
4.			
5.			
6.			
7.			
8.			
9.			

DEPARTMENT OF

Motor Vehicle Safety



Witness to certify that
Peter A. Fucano
Clifton City Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS Twelfth DAY OF June

TWO THOUSAND AND Nine

Joseph P. ...
SUPERINTENDENT
NEW JERSEY STATE POLICE

Gene ...
ATTORNEY GENERAL
STATE OF NEW JERSEY