

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWM-0025
Location: CLIFTON CITY POLICE
Calibration File No.: 02001 Calib. Date: 12/19/2014 Calib. No.: 00033
Certification File No.: 02002 Cert. Date: 12/19/2014 Cert. No.: 00022
Linearity File No.: 02003 Lin. Date: 12/19/2014 Lin. No.: 00022
Solution File No.: 02012 Soln. Date: 02/13/2015 Soln. No.: 00184
Sequential File No.: 02012 File Date: 02/13/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDXA S3-0033
Control Solution %: 0.100% Expires: 07/18/2015
Solution Control Lot: 13G115 Bottle No.: 0752

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:04S		
Control 1 EC	0.102%	17:05S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	17:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:06S		
Control 2 EC	0.101%	17:06S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	17:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:07S		
Control 3 EC	0.101%	17:08S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	17:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:09S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TURANO

First Name: PETER

MI: A

Signature: 

Badge No.: 204

Date: 02/13/2015



State of New Jersey

OFFICE OF THE ATTORNEY GENERAL
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
POST OFFICE BOX 7068
WEST TRENTON, NJ 08628-0068
(609) 882-2000

CHRIS CHRISTIE
Governor
KIM GUADAGNO
Lt. Governor

JOHN J. HOFFMAN
Acting Attorney General
COLONEL JOSEPH R. FUENTES
Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 7/29/2013

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 13G115

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1213 to 0.1220 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is July 18, 2015.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

[Handwritten signature of Ali M. Alaoui]

Ali M. Alaoui, Ph.D.
Research Scientist
NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 7th day of August, 2013.

[Handwritten signature of Linda L. W. Santos]
Notary




"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
Printed on Recycled Paper and Recyclable



DEPARTMENT OF

Motor Vehicle Safety



This is to certify that
Peter A. Furano
Clifton City Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO
CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS Twelfth DAY OF June

TWO THOUSAND AND Nine

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Law and Public Safety
 This is to certify that

Peter A. Turano
 Clinton City Police

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE **ALCO TEST 7110 MKIII-C**
 A METHOD TO DETERMINE INTOXICATION.
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THE 12th DAY OF June

TWO THOUSAND AND Nine

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	11/21/11	BCPA	R. J. W.
2.	11/13/13	S.C.C.S.T.A.	P. J. D.
3.			
4.			
5.			
6.			
7.			
8.			
9.			