

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
Location: CLIFTON CITY POLICE
Serial No.: ARWM-0025
Calibration File No.: 02053
Calib. Date: 06/17/2015
Calib. No.: 00034
Certification File No.: 02054
Cert. Date: 06/17/2015
Cert. No.: 00023
Linearity File No.: 02055
Lin. Date: 06/17/2015
Lin. No.: 00023
Solution File No.: 02075
Soln. Date: 08/21/2015
Soln. No.: 00190
Sequential File No.: 02075
File Date: 08/21/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 15D138
Model No.: CU-34
Serial No.: DDXA S3-0033
Expires: 04/13/2017
Bottle No.: 0026

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:56D		
Control 1 EC	0.099%	11:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:57D		
Control 2 EC	0.099%	11:58D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:58D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:59D		
Control 3 EC	0.100%	11:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:00D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: TURANO

First Name: PETER

MI: A

Signature: 


Badge No.: 204

Date: 08/21/2015

DEPARTMENT OF

**INDEPENDENT
LABORATORY**

STATE OF NEW JERSEY



Witness to certify that
Peter A. Fiano
Clifton City Police Department

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO
CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE Alcotest 7110 MKIII-C
A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY THIS Twelfth DAY OF June

TWO THOUSAND AND Nine

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY

DEPARTMENT OF
Traffic and Public Safety
 What is its motto? *What is its motto?*

Peter A. Unfrano
 CHIEF OF POLICE

IS QUALIFIED AND COMPETENT TO CONDUCT CHEMICAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF
 THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOHOL TEST 7110 MKIII-C**
 A METHOD TO DETERMINE INTOXICATION
 GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THIS 12TH DAY OF June

TWO THOUSAND AND Nine

[Signature]
 SUPERINTENDENT
 NEW JERSEY STATE POLICE

[Signature]
 ATTORNEY GENERAL
 STATE OF NEW JERSEY

ORIGINAL COURSE DATES

	DATE	Refreshers Course PLACE	INSTRUCTOR
1.	11/21/11	BC-PA	R. MW
2.	11/13/13	S.C.E.S.T.A.	R. D-7
3.			
4.			
5.			
6.			
7.			
8.			
9.			



State of New Jersey
 OFFICE OF THE ATTORNEY GENERAL
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
 POST OFFICE BOX 7068
 WEST TRENTON, NJ 08628-0068
 (609) 882-2000

CHRIS CHRISTIE
 Governor

JOHN J. HOFFMAN
 Acting Attorney General

KIM GUADAGNO
 Lt. Governor

COLONEL JOSEPH R. FUENTES
 Superintendent

CERTIFICATION OF ANALYSIS
0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.1174 to 0.1246 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 5/6/2015

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 15D138

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1203 to 0.1220 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-4.3, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is April 13, 2017.

As Research Scientist for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Ali M. Alaoui, Ph.D.
 Research Scientist
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 24th day of May, 2015.

Notary
MARY ELIZABETH MCLAUGHLIN
 ID # 2052190
 NOTARY PUBLIC
 STATE OF NEW JERSEY
 My Commission Expires Dec. 24, 2018



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer
 Printed on Recycled Paper and Recyclable

